



CSA INTERNATIONAL

Certificate of Compliance

Certificate: 1674552

Master Contract: 158394

Project: 1849753

Date Issued: 2007/01/03

Issued to: Honeywell Inc.
1100 Virginia Dr
Fort Washington, PA 19034-3260
USA
Attention: Edward Child

The products listed below are eligible to bear the CSA Mark shown



Issued by: Ezio Migliozi

Authorized by: Peter Schimmoeller, Manager
of Certification Services

PRODUCTS

CLASS 2258 02 - PROCESS CONTROL EQUIPMENT - For Hazardous Locations

Class I, Groups A, B, C and D; Class II, Groups E, F and G; Class III; Type 4

Thermal Conductivity gas Analyzers Type 7866, 32 Vdc max. input voltage, 30 W max power, two wire 4 to 20 mA or 0 to 20 mA output signals.

APPLICABLE REQUIREMENTS

- CSA Std C22.2 No 0.4-M1982 - Bonding and Grounding of Electrical Equipment
- CSA Std C22.2 No. 0.5-M1982 - Threaded Conduit Entries



Certificate: 1674552

Master Contract: 158394

Project: 1849753

Date Issued: 2007/01/03

- CSA Std C22.2 No. 25: 1992 - Enclosures for Use in Class II, Groups E, F and G – Hazardous Locations
- CSA Std C22.2 No. 30-M1986 - Explosion-Proof Enclosures for Use in Class I, Hazardous Locations
- CSA Std C22.2 No. 94-M91 - Special Purpose Enclosures Industrial Products
- CSA Std C22.2 No. 142-M1987 - Process Control Equipment

MARKINGS

- Submittor's identification;
- Model designation or equivalent
- Serial Number;
- Hazardous Location designation;
- Temperature code rating;
- Enclosure Type 4;
- Maximum pressure;
- The words "SEAL ALL CONDUITS WITHIN 18 INCHES"
- The words "DO NOT OPEN, EVEN WHEN ISOLATED, WHEN A FLAMMABLE/DUST ATMOSPHERE IS PRESENT";
- The warning: "DO NOT OPEN WHILE ENERGIZED"